

# OK!

## OK! SP


### Technical Data



### Ear simulator data

EN / IEC 60118 and IEC 60711

#### Output sound pressure level




SuperPower BTE with battery size 13 and omnidirectional microphone.

#### Warning to hearing care professionals:

This hearing instrument has an output sound pressure level that can exceed 132 dB SPL. Special care should be taken when fitting this instrument as there is a risk of impairing the residual hearing of the user.

#### Acoustic gain



Unless otherwise specified, all data obtained are measured with the hook type HE7.

Note: Measurements with pure tones of a digital hearing instrument can result in a wavy frequency response. This is an artifact resulting from the use of a narrowband input signal and does not reflect the actual performance with naturally occurring broadband input signals.

Frequency range	<100 Hz - 6300 Hz		
Total harmonic distortion	500 Hz	800 Hz	1600 Hz
	5.0%	4.0%	2.0%
Battery current	Quiescent	Working	
	1.1 mA	1.2 mA	
Equivalent input noise level	20 dB SPL		


#### Dynamic data

Compression	Attack time	Recovery time
	1 ms	20 ms


## 2cm<sup>3</sup> coupler data

ANSI S3.22-1996


### Output sound pressure level



### Input / Output characteristics at 2000 Hz



### Acoustic gain



Frequency range	<100 Hz – 6200 Hz		
Total harmonic distortion	500 Hz	800 Hz	1600 Hz
	4.0%	2.0%	1.0%
Equivalent input noise level	21 dB SPL		

### Dynamic data

Compression	Attack time	Recovery time
	1 ms	40 ms